

CERTIFICATE OF ANALYSIS

Order Type: Medical Cannabis
Order ID: OR2025-7979

Customer ID: 1000
Customer Name: BingerLabs

Harvest/Extract Lot: None
Harvest/Extract Batch: None

Cultivar (Strain): 40mg CBDpatch
Sample Date: 11/20/2025

Lab ID: SA2025-1875
Date Received: 11/20/2025
Remarks:

Sample Matrix: Water
Date Completed: 11/22/2025

CANNABINOID (POTENCY) PROFILE

Analysis Date/Time: 11/20/2025 1515
Analyst: OL

Method: 2025-11-18T20:36:50Z
Instrument: Agilent 1100

Moisture Content (%): -
Water Activity (aw): -


Cannabinoid	Result (%)	Result (mg/g)	Reporting Limit (mg/g)	Result (mg/mL)	Per Unit (mg)
CBD	4.42	44.2	0.073	-	41
CBDa	-	-	0.073	-	-
CBDv	0.021	0.21	0.073	-	0.168
Δ9-THC	-	-	0.073	-	-
Δ8-THC	-	-	0.073	-	-
THCa	-	-	0.073	-	-
THCv	-	-	0.073	-	-
CBC	-	-	0.073	-	-
CBG	-	-	0.073	-	-
CBGa	-	-	0.073	-	-
CBN	-	-	0.073	-	-
TOTAL	4.44	44			42
TOTAL THC	-	-			-
TOTAL CBD	4.42	44			41



UNIT MASS (g): 0.80

"-" Not detected above RL.

Cannabinoid Distribution (% of Total Cannabinoids)



Deviations from standard operating procedure: None

Recoveries for all analyte standards: 90-110%
Replicate Uncertainties: <5% RSD, <20% RPD
Sample/Reagent Blanks: <RL for all analytes

Values for plant matter are adjusted for moisture content.

Total THC = (THCa x 0.877) + Δ9-THC
Total CBD = (CBDa x 0.877) + CBD

Percentage results are reported by mass.
mg/g results are reported as mass component per mass material.

Abbreviations: UV - Ultraviolet, HPLC - High Pressure Liquid Chromatography, RL - Reporting Limit, RPD - Relative Percent Difference, RSD - Relative Standard Deviation

This information is provided as a service and makes no claims of efficacy and/or safety of this product. Results are applicable only for the sample(s) analyzed and for the specific analysis conducted. This report is for informational purposes only and should not be used to diagnose, treat, or prevent any medical-related symptoms.

The statements and results herein have not been approved and/or endorsed by the FDA.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Felling Analytical Services and Technology (F.A.S.T.), LLC

Kyle W. Felling
Kyle W. Felling, Ph.D.
Laboratory Director